



T09 Stair Suitability

A= Angle – 35° maximum.

L = Landing – Top 107cm minimum.*

Landing – Bottom 97cm minimum.*

G= Going – (angle dependant, e.g. 26cm for 18cm rise)

R= Rise – 8 cm minimum.

18.5cm maximum.

W= Width – 82 cm minimum.*

Compliance - BS 585-1:1989 Section K1:–

Straight “Institutional and assembly” stairs – width permitting.

All other stairs must be considered on a case by case basis.

Not suitable for tapered, spiral or helical stairs.

* Based on T09 with a Wheeltech Enigma self propel 18” wheelchair. Figures will vary with other wheelchairs.